



Introducing neurafit[™]

Our solution to preventing spinal misconnections

Our ISO standard NRFit™ connectors minimise the risk of medication being delivered to the patient via the wrong route

The problem

When used for neuraxial and regional anaesthesia, universal Luer connections have led to administrative errors¹, which in some cases has proven fatal².

Problems occurred when intravenous (IV) medicines were administered by the spinal route and vice versa; when anaesthetics were delivered through IV lines³.

The solution

It is widely recognised that the use of a non-Luer connection can limit these errors⁴ which has generated an increasing need for better connections⁵. The result is the ISO 80369-6 standard named NRFit⁶. In line with these new standards specifying small-bore connectors for neuraxial and regional anesthesia, Vygon has developed neurafit.

NRFit compliant

Featuring NRFit standard small-bore connectors, the benefits of neurafit include:

- Avoids any connection to Luer devices
- A one-piece reversed design, and a smaller diameter, to minimise dead space
- No misconnection.





neurafit™: the advantages

Made to be sturdy and exceptionally strong from premium-quality steel, neurafit is available with two different needle designs to meet specific clinical needs. The device delivers a host of safety benefits and efficiency features.



Quincke needle design

Ideal for performing an effective lumbar puncture with a raid return of cerebrospinal fluid (CSF).



Whitacre pencil point design

Reduces the risk of post spinal headache during intrathecal injection. The shape, size and positioning of the eye ensures optimum needle placement and excellent CSF response time.



Safe

- Easy identification of our neura**fit** range with yellow protectors; the colour associated with NRFit standards
- Recessed notch indicates the direction of the needle eye plus accurate administration of anaesthetic.



Efficient

- Crystal clear hub has fast CSF flashback
- Easy to see CSF through magnified hub
- Spinal introducer with precision matched stylet delivers quick and easy one step removal. Plus instant confirmation of needle tip position.



Ease of use

- Available with slip and lock connections
- Familiar connections mean there is no need to change established habits when switching to NRFit devices
- Introducer has an ergonomic hub for easy handling.



Whitacre needles

Vygon Code	NHSSC	Description	Colour	Needle		Introducer		Quantity
				Diameter (G)	Length (mm)	Diameter (G)	Length (mm)	Box/case
000181042A1	FTR2422	neurafit whitacre needle		27	90	22	38	25/1000
000181942A1	FTR2421	neurafit whitacre needle	•	26	90	20	38	25/1000
000181052A1	FTR2419	neurafit whitacre needle	•	25	90	20	38	25/1000
000181252A1	FTR2420	neurafit whitacre needle	•	25	120	20	38	25/1000
000181072A1	FTR2417	neurafit whitacre needle	•	22	90	18	38	25/1000
00518127A1	FTR2418	neurafit whitacre needle	•	22	120	-	-	25/1000

Quincke needles

Vygon Code	NHSSC	Description	Colour	Needle		Introducer		Overstitus
				Diameter (G)	Length (mm)	Diameter (G)	Length (mm)	Quantity Box/case
00518005A1	FTR2429	neura fit quincke needle		25	90	-	-	25/1000
000180052A1	FTR2430	neurafit quincke needle		25	90	20	38	25/1000
00518347A1	FTR2426	neurafit quincke needle	•	22	38	-	-	25/1000
00518397A1	FTR2427	neurafit quincke needle	•	22	90	-	-	25/1000
00518027A1	FTR2428	neurafit quincke needle		22	120	-	-	25/1000
00518349A1	FTR2424	neurafit quincke needle		20	38	-	-	25/1000
00518399A1	FTR2425	neurafit quincke needle		20	90	-	-	25/1000
00518392A1	FTR2423	neura fit quincke needle		18	90	-	-	25/1000

References

- 1 Milamed D, Brown K, Murphy E. Luer's lure: from an international standards perspective. Anesthesiology 2012;117:1358–1363.
- 2 Veisi F, Salimi B, Mohseni G, Golfam P, Kolyaei A. Accidental intrathecal injection of tranexemic acid in cesarean section: a fatal medication error. APSF Newslett 2010;25:9.
- 3 Smetzer J, Baker C, Byrne F, Cohen M. Shaping systems for better behavioral choices: lessons learned from a fatal medication error. Jt Comm J Qual Patient Saf 2010;36:152–163.
- 4 Patel S, Loveridge R. Obstetric neuraxial drug administration errors: a quantitative and qualitative analytical review. Anesth Analg 2015;121:1570–1577.
- 5 Birnbach D, Brull S, Prielipp R. J'accuse: failure to prevent epidural and spinal catheter misconnections. A A Case Rep 2016;6:107–110.
- 6 https://www.iso.org/standard/50734.html

For further information, please contact: vygon@vygon.co.uk

The specifications shown in this leaflet are for information only and are not, under any circumstances, of a contractual nature.

NRFit™ is a trademark of GEDSA and is used with their permission.

Vygon (UK) Ltd, The Pierre Simonet Building, V Park, Gateway North, Latham Road, Swindon, Wiltshire SN25 4DL

